

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

FEB 12 2010

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130115</u>
Name <u>DOUG E. HARRIS</u>		ID# <u>5530</u> Date <u>FEB 7, 2010</u>
A	Agency <u>BRISTOL BAY POLICE</u>	Phone # <u>246-4222</u>
Instrument Location <u>BBPD KING SALMON AK</u>		
Alco S/N <u>X124623</u>		Target Value <u>.076</u> High Pressure _____
B	Alco Test Values <u>.075</u> <u>.077</u> 1 st Alco <u>2-12-10</u> 2 nd Alco	
Signature <u>Doug E. Harris</u>		
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of February, 2010.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

VERIFICATION OF CALIBRATION REPORT

of DataMaster edm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster edm S/N

130115
2/22/10
X124623

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER edm 130115

FEBRUARY 07, 2010

SUBJECT'S NAME:
UCR015 0016 E

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.I. #: 0000001

DEPT/AGENCY: NKN1

CASE/REPORT: TEST

TEST TYPE: U

ALCO TARGET VALUE: 076

ALCO S/N: X124623

BREATH ANALYSIS

.076 ADJUSTED FOR 29.83 in	.075	11:42
ALCO TARGET	.000	11:43
BLANK TEST	.000	11:44
INTERNAL STANDARD	VERIFIED	11:45
ALCO TO 29.83 in	.077	11:46
BLANK TEST	.000	11:47
SUBJECT SAMPLE	.000	11:48
BLANK TEST	.000	11:49
ALCO TO 29.83 in	.077	11:50
BLANK TEST	.000	11:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER edm 130115

FEBRUARY 07, 2010

TIME 11:57

DIAGNOSTIC CHECK

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.83 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

ALCO (14) :- /0123456789:;(<=>?@ACDEF6
GHIJKLPQRSTUWXYZ:~! " abcdefghijklmno
pqrstuvwxyz{|}~